

## Manufacturer's Certificate of Validation

Certificate No.	Calibratoin Date	Page
I19060608	April 26th, 2019	1 of 1

PLUSTRACE TECHNOLOGIES certifies that the products mentioned on the following page(s) have been thoroughly tested, validated and met performance accuracy specifications over the stated ranges.

### Reference Instrument

Device	Model	Calibration Date	Accuracy/Volatility	Serial Number
FLUKE Standard Temperature Instrument	1551A	August 28,2018	±0.05 °C	FQDT-378

All reference instruments are periodically calibrated by HNIM, a CNAS (China National Accreditation Service) accredited laboratory and traceable to NIST according to Multi-Recognition Agreement. CNAS is full member of the International Laboratory Accreditation Cooperation. This Validation Certificate is also accepted by all signatories to the ILAC MRA.

### Environment Conditions

<b>Air Temperature</b>
20.5 °C
<b>Relative Humidity</b>
53.0%

### Valiadiion Information

Validation Points	Tolerance	Result
+30 °C	±0.5 °C	PASS
0 °C	±0.5 °C	PASS
-10 °C	±0.5 °C	PASS

### Product Information

Product Name	Serial Number	Production Date	Expiry Date	Quantity
Plustrace I1	I19008001-I19008020	06/2019	06/2021	20

This certificate was automatically generated and valid for 2(two) years from the date of production.

Approved by: 

